

MAY 08 2014

5. 510(k) Summary

Device Trade Name: STALIF C®

Manufacturer: Centinel Spine, Inc.
900 Airport Road, Suite 3B
West Chester, PA 19380

Contact: Mr. John Parry
Development Manager
Phone: (484) 887.8813
J.Parry@centinelspine.com

Prepared by: Justin Eggleton
Musculoskeletal Clinical Regulatory Advisers, LLC
1331 H Street NW, 12th Floor
Washington, DC 20005
Phone: (202) 552-5800
Fax: (202) 552-5798

Date Prepared: October 15, 2013

Classifications: 21 CFR §888.3080, Intervertebral body fusion device

Class: II

Product Codes: OVE

Indications For Use:

The STALIF C® is intended to be used as an intervertebral body fusion cage as a standalone system used with bone screws provided and requires no additional supplementary fixation systems. It is inserted between the vertebral bodies into the disc space from levels C2 to T1 for the treatment of cervical degenerative disc disease (defined as neck pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies). The device system is designed for use with autograft bone to facilitate fusion. STALIF C® is intended to be used at one level.

The cervical cage is to be used in a skeletally mature patient who has had six weeks of non-operative treatment prior to implantation of the cage.

Device Description:

STALIF C® is a radiolucent intervertebral body fusion cage with unicortical cancellous bone screws. It is intended to be used as an IBF cage without supplementary fixation. The cross section profile of the STALIF C® is similar to that of the vertebral body endplate

with central cavity that can be packed with autograft. STALIF C® is manufactured from PEEK-OPTIMA® LTI with titanium alloy screws and X-ray marker wires manufactured from unalloyed Tantalum (ASTM F-560). The purpose of this 510(k) was to include a titanium coating (ASTM F1580) on both endplates.

Predicate Device:

The subject STALIF C® intervertebral body fusion device is substantially equivalent to predicate STALIF C™ (K120819, K072415) with respect to indications, design, function, and materials.

Substantial Equivalence:

Testing performed indicate that the STALIF C® is as mechanically sound as predicate devices. Testing included static compression, static torsion, static compression-shear, dynamic compression, dynamic torsion, dynamic compression-shear, expulsion, and subsidence per ASTM F2077 and F2267. The results demonstrate that the acceptance criteria defined in the Design Control Activities Summary were met.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

May 8, 2014

Centinel Spine, Incorporated
% Musculoskeletal Clinical Regulatory Advisers, LLC
Mr. Justin Eggleton
Director, Spine Regulatory Affairs
1331 H Street Northwest, 12th Floor
Washington, District of Columbia 20005

Re: K133200

Trade/Device Name: STALIF C®
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral body fusion device
Regulatory Class: Class II
Product Code: OVE
Dated: April 7, 2014
Received: April 8, 2014

Dear Mr. Eggleton:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Ronald P. Jean -S for

Mark N. Melkerson
Division Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

4. Indications for Use

510(k) Number (if known): K133200

Device Name: STALIF C®

The STALIF C® is intended to be used as an intervertebral body fusion cage as a standalone system used with bone screws provided and requires no additional supplementary fixation systems. It is inserted between the vertebral bodies into the disc space from levels C2 to T1 for the treatment of cervical degenerative disc disease (defined as neck pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies). The device system is designed for use with autograft bone to facilitate fusion. STALIF C® is intended to be used at one level.

The cervical cage is to be used in a skeletally mature patient who has had six weeks of non-operative treatment prior to implantation of the cage.

Prescription Use √
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE
IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Ronald P. Jean -S

(Division Sign-Off)

Division of Orthopedic Devices

510(k) Number: K133200